



Alcohol Breathalysers

Saving Lives! Saving Licences!

Alcohol Breathalysers (Pty) Ltd
Co. Reg. No. 2016/345956/07
VAT Reg. No. 4510217161
29 Plantation Road
The Gardens, Norwood
Johannesburg, 2192, RSA
Johannesburg: +27 10 140 0785
Cape Town: +27 21 007 4476
Durban: +27 31 003 8917
sales@breathalysers.co.za
<https://breathalysers.co.za/>

BREA/CL/SH/1

Specifications

Model: iBlow10 Basic Pack

Type: No Contact Industrial & Law Enforcement
Breathalyser Screener

Contents:

- iBlow10 device X 1pc
- Blow caps X 3pcs
- Hand-strap X 1pc
- User manual X 1pc
- USB cable X 1pc for data withdrawal via max30 software **(Not Supplied for Re-Charging)**



		Details
Sensor Type		SuraCell Sensor (Fuel Cell Sensor)
Battery /Power Supply		6 x AA Size alkaline battery / rechargeable battery (Non-Manganese)
Weight		1kgs
Size (mm)		30cm X 10cm X 10cm
Test Result	LED Alert	Zero: Green / Low: Red / Hi: Blinking Red
	FND Display	0.00/LO/HI or numerical result (Display "HI" in numerical result mode means more than 1.00 g/l of BAC)
	Buzzer Alarm	Zero: 1 Beep sound Low: Repeating short beep sound High: Long Beep
Warming Up Time		Within 10 seconds
Time for Analysis		Within 2 seconds (at 0.00 BAC)
Maximum Test		12 Times / 1 Minute
Working Temperature		-5°C ~ 55°C (optional -10°C ~ -55°C)
Storage		-10°C ~ 60°C
Mouthpieces		Blow cap (Replaceable & washable passive cone type)
Auto Power Off		Adjustable 30 to 900 seconds
Duration Time for Display		Adjustable 1 to 9 seconds
Calibration period		Default – Every 65,000 tests or 12 months whichever comes first (Adjustable by Local Services Centre Agent)
Additional Features		Built-in Magnet
		Reminder for calibration & grace period
		Calibration protection by Webserver Connecting
		Manual LED lighting: White, Red blinking, Red, Red & blue

Packaging



Breathalyser in use

