



Alcohol Breathalysers

Saving Lives! Saving Licences!

Alcohol Breathalysers (Pty) Ltd
Co. Reg. No. 2016/345956/07
VAT Reg. No. 4510217161
29 Plantation Road
The Gardens, Norwood
Johannesburg, 2192, RSA
Johannesburg: + 27 10 140 0785
Cape Town: +27 21 007 4476
Durban: +27 31 003 8917
sales@breathalysers.co.za
<https://breathalysers.co.za/>

BREA/CL/SH/7

Specification

Model: Alcoscan ALP-1 Full Pack

Type: ALP1 BT for Industry and
Law-Enforcement
Breathalyser – Full Pack

Contents:

- ALP-1 device X 1pc
- Mouthpieces X 5pcs
- Hand-strap X 1pc
- Max30 Software X 1pc
- USB cable X 1pc for downloading data via max30 software (Not Supplied for Re-Charging)
- User manual X 1pc
- Batteries X 2pcs
- Leather pouch X 1pc
- 8gig SD Card X 1pc
- Blow Cup X 1pc
- Bluetooth Feature



		Details
Sensor Type		Fuel Cell Sensor
Battery / Power Supply		2 x AA size Alkaline Battery / Rechargeable Batteries (Non- Manganese)
Weight		198g (including AA Alkaline Batteries / 500g when packed)
Size(mm)		133mm X 64mm X 33mm
Test Result	LED Alert	Zero: Green / Low: Red / Hi: Blinking Red
	FND Display	Fast Mode: Pass – Green / Fail – Red
		Normal Mode: 0.000%BAC – Green / 0.001%BAC – Red
	Buzzer Alarm	Zero: 1 Beep sound Low: Repeating short beep sound High: Long Beep
Accuracy		+/- 0.005 %BAC at 0.100 %BAC
Warming Up Time		Within 5 seconds to 2 minutes
Response Time		Within 10 seconds at 0.100%BAC

Working Temperature	-5°C ~ 40°C (optional -10°C ~ -55°C)
Storage	500 Tests on Memory (Memory Extension Via SD Card Slot)
Mouthpieces	5 x Mouthpieces
Auto Power Off	15 Seconds
Duration Time for Display	Adjustable 5 to 10 seconds
Calibration period	Default – Every 15,000 tests or 12 months whichever comes first (Adjustable by Local Services Centre Agent)
Additional Features	Bluetooth Functionality
	By-Annual calibration reminder
	Calibration protection by administrator Pin
	Compatible with the iSober App Free / Enterprise Service
	Certified Water resistant to IP54 rating IEC 60529: 2001
	PASS/FAIL Fast testing set to provide a Fail result from 0.01% B.A.C

Packaging



Breathalysers In Use

